

# Installation Guide Philips Avalon CTS Mounting Adapter


#### **Parts Reference**

The following parts and hardware are included in this installation kit (hardware not shown).


Item #	Description	Qty
1	Avalon Mounting Adapter	1
2	M3.5 x 6mm Pan Head Machine Screw (PHMS)	4

## **Tools Required:**


Phillips Screwdriver (not supplied)




### **Installing the Philips Avalon CTS Mounting Adapter**



1. Using a Phillips screwdriver, attach the mounting adapter to the Avalon with (4) M3.5 x 6mm PHMS.



2. Pull the plunger pin at the front of the mounting plate. Slide the mounting adapter into the mounting plate until the plunger pin snaps into the adapter plate clearance hole.



**3.** Tighten the (2) Nylon screws located towards the front of the mounting plate to secure.

### **Routine Maintenance**

Periodically inspect all fasteners associated with the mounting system. Tighten as necessary for optimal operation and safety.

## **Cleaning the Mounting Assembly**

- 1. The mounting assembly may be cleaned with most mild, non-abrasive solutions commonly used in the hospital environment (e.g. diluted bleach, ammonia, or alcohol solutions).
- 2. The surface finish will be permanently damaged by strong chemicals and solvents such as acetone and trichloroethylene.
- 3. Do not use steel wool or other abrasive material to clean the mounting assembly.
- **4.** Damage caused by the use of unapproved substances or processes will not be covered by warranty. We recommend that you test any cleaning solution on a small area of the mounting assembly that is not visible to verify compatibility.
- 5. Never submerge or allow liquids to enter the mounting assemblies. Wipe any cleaning agents off the mounting assembly immediately using a water-dampened cloth. Dry all mounting assemblies thoroughly after cleaning.

**CAUTION:** GCX makes no claims regarding the efficacy of the listed chemicals or processes as a means for controlling infection. Consult your hospital's infection control officer or epidemiologist. To clean or sterilize mounted instruments or accessory equipment, refer to the specific instructions delivered with those products.